

FAQ

INNO-LiPA HPV



- [ON WHAT KIND OF SAMPLES CAN WE USE THE INNO-LiPA™ HPV GENOTYPING EXTRA KIT?](#)

The test has been validated on cervical cells collected in Surepath® medium. But it can also be used on other kind of samples, those have not been validated though. Articles are available describing use of other sample types. The use of other sample types with INNO-LiPA™ HPV Genotyping Extra should be validated in your laboratory.

- [DO YOU ALSO PROVIDE THE COLLECTION KIT FOR THE CERVICAL SAMPLES?](#)

Fujirebio Europe N.V. only provides the INNO-LiPA™ HPV Genotyping Extra and Amplification kit, no collection kits.

- [HOW LONG CAN WE STORE THE ORIGINAL CERVICAL SAMPLES IN LIQUID MEDIUM BEFORE TESTING?](#)

Storage conditions should be checked with the manufacturer's instructions. Validation of long term sample stability was not performed.

- [IS THE INNO-LiPA™ HPV GENOTYPING EXTRA AMPLIFICATION KIT COMPATIBLE WITH OTHER HPV GENOTYPING KITS?](#)

The Amplification kit has been validated together with the INNO-LiPA™ Genotyping Extra kit. We have no data on the compatibility with other HPV genotyping kits.

- [WHY SHOULD I USE YOUR TEST?](#)

Due to the small SPF10 primers in our kit (65bp), the test is highly applicable with all kind of HPV positive samples.